

CERTIFICATE OF ANALYSIS

Product name: Chemical indicator for steam (Type 4)

Model: UGCI0104

Lot: 2025-07

Manufacture Date: 22-07-2022

Expiry Date: 22-07-2025

Performance Specifications:

Complete color change to black	15.1min at 121°C	3.3 min at 134°C
---------------------------------------	-------------------------	-------------------------

This document certifies that the product listed above, chemical indicator does not contain toxic substances in quantities harmful to the health of individuals before, during, or after exposure to sterilization processes.

This also certifies that the product listed above satisfies performance requirements after exposure to dry saturated steam when tested under the conditions described in ISO 11140-1 & GB 18282-1. This chemical indicator is classified as a Type 4 indicator as specified by ISO 11140-1 & GB 18282-1.

These products are stable after exposure to steam and may be kept as a permanent record.


Technical Director

Ugaiya Bio-Sciences Co.,Ltd

Add:Nishifunahashi 2-47-11, Hirakata City, Osaka, Japan 573-1122